

## Instrument Response Information

Broadband Sensor	Reftek 151 120 A
Sensitivity (V/m/s)	2000
Natural Period	120 Sec
Normalisation Factor	65135
Normalisation Frequency	1 Hz
No. of Zeros	2
No. of Poles	4
Poles and Zeros	
Zeros (rad/sec)	0
	0
Poles (rad/sec)	$-0.03702 + 0.03702 j$
	$-0.03702 - 0.03702 j$
	$-177.72 + 177.72j$
	$-177.72 - 177.72j$