

Instrument Response Information

Broadband Sensor	Reftek 151B-120
Sensitivity (V/m/s)	2000
Natural Period	120 Sec
Normalisation Factor	98696
Normalisation Frequency	1 Hz
No. of Zeros	2
No. of Poles	4
Poles and Zeros	
Zeros (rad/sec)	0
	0
Poles (rad/sec)	$-0.036788 + 0.037216 j$
	$-0.036788 - 0.037216 j$
	$-222.144 + 222.144j$
	$-222.144 - 222.144j$